



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

South Texas Calibration Labs / Fernando Damián Casiano Flores
Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Dimensional, Electrical, Thermodynamic, Chemical, Mechanical, Mass, Force & Weighing Device, Time & Frequency, and Acoustic Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Initial Accreditation Date:

Issue Date:

Expiration Date:

April 10, 2010

December 18, 2022

January 31, 2025

Accreditation No.:

Certificate No.:

66293

L22-871-1

Tracy Szerszen
President

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjllabs.com

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Page 1 of 20



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Micrometer (Inside, Outside, Depth) ^{FO}	0.05 in to 42 in (1.27 mm to 1 066.8 mm)	(385 + 10L) μ in [(9.78 + 10 x 10 ⁻³ L) mm]	Gage Blocks ASME B89.1.13
Dial Indicator ^{FO}	0.05 in to 1 in (1.27 mm to 25.4 mm)	(112 + 9L) μ in (2.84 + 9 x 10 ⁻³ L) mm	Gage Blocks ASME B89.1.10M
Digital Indicator ^{FO}	0.05 in to 1 in (1.27 mm to 2.54 mm)	380 μ in (9.7 μ m)	Gage Blocks ASME 89.1.10M
Caliper ^{FO}	0.05 in to 42 in (1.27 mm to 1 066.8 mm)	(483 + 14L) μ in (12.27 + 14 x 10 ⁻³ L) mm	Gage Blocks ISO 6906 ISO 3599 JIS B 7507
Surface Plate Flatness Only ^O	8.5 in to 24 in Diagonal (215.9 mm to 609.6 mm)	120 μ in (3.048 μ m)	CMM and Planikator ASME B89.7.3
Surface Plate Repeat Measurement Only ^O	0.002 in (0.050 8 mm)	40 μ in (1.01 μ m)	Repeat o Meter GGG-P-453c
Thread Plug Gage (Pitch Diameter) ^{FO}	0 - 80 to 4 -12	150 μ in (3.81 μ m)	Measuring Wire Set FED-STD-H28
Adjustable Thread Ring Gage (Pitch Diameter) ^F	0-80 to 4-12	210 μ in (4.2 μ m)	Super Micrometer Trimos-Sylvac 80 ANSI/ASME B1.16M ANSI/ASME B1.2
Adjustable Thread Ring Gage (Minor Diameter) ^F	4 in Maximum (100 mm Maximum)	64 μ in (1.6 μ m)	Super Micrometer Trimos-Sylvac 80 ANSI/ASME B1.16M ANSI/ASME B1.2
Height Gage ^{FO}	0.05 in to 42 in (1.27 mm to 1 066.8 mm)	(245 + 21L) μ in (6.22 + 21 x 10 ⁻³ L) mm	Gage Blocks JIS B 7517
Coating Thickness Gage Ferrous ^F	52.47 μ m to 179.13 μ m	1.2 μ m	Coating Thickness Films
Coating Thickness Gage Non-Ferrous ^F	52.44 μ m to 175.66 μ m	1.1 μ m	Internal Procedure: STC-PC-DIM-016
CMM –Verification Volumetric	Up to 30 in	77 μ in	Renishaw MCG ASME B89.4.10360.2 ASME B89.4.1
CMM –Verification Volumetric	0.5 to 20 in	(30.34 + 1.34L) μ in	Gage Block Gages ASME B89.4.10360.2 ASME B89.4.1
Gage Blocks ^F	0.05 in to 4 in (0.5 mm to 101.6 mm)	(2.6 + 1L) μ in [(0.08 + 1 x 10 ⁻³ L) μ m]	Gage Block Comparison ASME B89.1.9
Protractor ^{FO}	0° to 90°	0.01°	Angle Blocks Internal Procedure: STC-PC-DIM-018



Certificate of Accreditation: Supplement
South Texas Calibration Labs /
Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
 Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
 Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Video Measuring Machine ^O	0.04 in to 12 in (1.01 mm to 304.8 mm)	220 μ m (5.51 μ m)	Glass Standard Internal Procedure: STC-PC-DIM-031
Angle Block ^F	0° to 180°	0.03°	CMM Internal Procedure: STC-PC-DIM-017
Measure Tape ^F	1 in to 300 in (25.4 mm to 7 620 mm)	0.007 1 in 1.8 mm	Video Measuring Machine Internal Procedure STC-PC-DIM-024
Rules ^F	0.02 in to 48 in (0.5 mm to 1 220 mm)	11 μ m (0.28 mm)	Standard Rule & Video Measuring Machine Internal Procedure STC-PC-DIM-026
Roughness Tester Ra (Fixed point) ^{FO}	117 μ m	0.76 μ m	Roughness Standard EAL-G20 ASME B46.1
Level ^F	2 in to 8 in	320 μ m	Precision Lever Internal Procedure: STC-PC-DIM-023
Optical Comparator ^O X axis Linearity	12 in Maximum (304.8 mm) Maximum	(116.89 + 2.92L) μ m	Glass Standard Internal Procedure: STC-PC-DIM-007
Y axis Linearity	12 in Maximum (304.8 mm) Maximum	(2.95 + 2.92 x 10 ⁻³ L) μ m	
Optical Comparator Angularity ^O	0° to 180°	0.02°	
Optical Comparator Magnification ^O	10X	0.03 %	
	20X	0.07 %	
	50X	0.08 %	
Cylindrical Diameter Inside (Plain Ring Gage) ^{FO}	4 in Maximum (100 mm Maximum)	30 μ m (0.77 μ m)	Super Micrometer Trimos-Sylvac 80 ASME B89.1.6 ASME B89.1.5
Cylindrical Diameter Outside (Pin Gages, Plain Plug Gage, Cylindrical Gages) ^{FO}	4 in Maximum (100 mm Maximum)	37 μ m (0.94 μ m)	

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output DC Voltage ^{FO}	Up to 330 mV	0.006 % of reading + 3 μ V	Fluke 5500A Internal Procedure STC-PC-ELE-006



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output DC Voltage ^{FO}	Up to 3.3 V	0.005 % of reading + 5 μ V	Fluke 5500A Internal Procedure STC-PC-ELE-006
	Up to 33 V	0.005 % of reading + 50 μ V	
	30 V to 330 V	0.005 5 % of reading + 500 μ V	
	100 V to 1 000 V	0.005 5 % of reading + 1 500 μ V	
	1 000 V to 5 000 V	2 % of reading	Hipot 3565D Internal Procedure STC-PC-ELE-004
	5 000 V to 25 000 V	2.5 % of reading	Hipot HD 125 Internal Procedure STC-PC-ELE-004
Equipment to Measure DC Voltage ^{FO}	Up to 100 mV	0.000 9 % of reading + 0.3 μ V	HP 3458A Internal Procedure STC-PC-ELE-006
	100 mV to 1V	0.000 8 % of reading + 0.3 μ V	
	1 V to 10 V	0.000 8 % of reading + 0.5 μ V	
	10 V to 100 V	0.001 % of reading + 30 μ V	
	100 V to 1 000 V	0.001 % of reading + 0.1 mV	
	Up to 20 kV	0.04 % of reading + 4 V	Vitretek 4600 Internal Procedure STC-PC-ELE-006
Equipment to Output DC Current ^{FO}	Up to 3.3 mA	0.013 % of reading + 0.05 μ A	Fluke 5500A Internal Procedure STC-PC-ELE-006
	Up to 33 mA	0.01 % of reading + 0.25 μ A	
	Up to 330 mA	0.01 % of reading + 3.3 μ A	
	Up to 2.2 mA	0.03 % of reading + 44 μ A	
	Up to 11 A	0.06 % of reading + 330 μ A	
	11 A to 500 A	0.5 % of output + 0.5 A	Fluke 5500A / 50 Turn Coil Internal Procedure STC-PC-ELE-009
Equipment to Measure DC Current ^{FO}	Up to 1 μ A	0.002 % of reading	HP 3458A Internal Procedure STC-PC-ELE-006
	1 μ A to 10 μ A	0.002 % of reading	
	10 μ A to 100 μ A	0.002 % of reading	
	100 μ A to 1 mA	0.002 % of reading	
	1 mA to 10 mA	0.002 % of reading	
	10 mA to 100 mA	0.003 5 % of reading	
	100 mA to 1 A	0.011 % of reading	
	1 A to 500A	0.06 % of reading	HP 34401A & Shunt (100 A/ 100 mV) Internal Procedure STC-PC-ELE-006



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output AC Voltage At the listed frequencies ^{FO}			Fluke 5500A Internal Procedure STC-PC-ELE-006
10 Hz to 45 Hz	1 mV to 32.999 mV	0.35 % of reading + 20 μ V	
45 Hz to 10 kHz	1 mV to 32.999 mV	0.15 % of reading + 20 μ V	
10 kHz to 20 kHz	1 mV to 32.999 mV	0.2 % of reading + 20 μ V	
20 kHz to 50 kHz	1 mV to 32.999 mV	0.25 % of reading + 20 μ V	
50 kHz to 100 kHz	1 mV to 32.999 mV	0.35 % of reading + 33 μ V	
100 kHz to 500 kHz	1 mV to 32.999 mV	1 % of reading + 60 μ V	
Equipment to Output AC Voltage At the listed frequencies ^{FO}			
10 Hz to 45 Hz	33 mV to 329.999 mV	0.25 % of reading + 50 μ V	
45 Hz to 10 kHz	33 mV to 329.999 mV	0.05 % of reading + 20 μ V	
10 kHz to 20 kHz	33 mV to 329.999 mV	0.1 % of reading + 20 μ V	
20 kHz to 50 kHz	33 mV to 329.999 mV	0.16 % of reading + 40 μ V	
50 kHz to 100 kHz	33 mV to 329.999 mV	0.24 % of reading + 170 μ V	
100 kHz to 500 kHz	33 mV to 329.999 mV	0.7 % of reading + 330 μ V	
Equipment to Output AC Voltage At the listed frequencies ^{FO}			
10 Hz to 45 Hz	0.33 V to 3.29 999 V	0.15 % of reading + 250 μ V	
45 Hz to 10 kHz	0.33 V to 3.29 999 V	0.03 % of reading + 60 μ V	
10 kHz to 20 kHz	0.33 V to 3.29 999 V	0.08 % of reading + 60 μ V	
20 kHz to 50 kHz	0.33 V to 3.29 999 V	0.14 % of reading + 300 μ V	
50 kHz to 100 kHz	0.33 V to 3.29 999 V	0.24 % of reading + 1 700 μ V	
Equipment to Output AC Voltage At the listed frequencies ^{FO}			
10 Hz to 45 Hz	3.3 V to 32.999 V	0.15 % of reading + 2 500 μ V	
45 Hz to 10 kHz	3.3 V to 32.999 V	0.04 % of reading + 600 μ V	
10 kHz to 20 kHz	3.3 V to 32.999 V	0.08 % of reading + 2 600 μ V	
20 kHz to 50 kHz	3.3 V to 32.999 V	0.19 % of reading + 5 000 μ V	
50 kHz to 100 kHz	3.3 V to 32.999 V	0.24 % of reading + 17 000 μ V	



Certificate of Accreditation: Supplement
South Texas Calibration Labs /
Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
 Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
 Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output AC Voltage At the listed frequencies ^{FO}			Fluke 5500A Internal Procedure STC-PC-ELE-006
45 Hz to 1 kHz	33 V to 329.999 V	0.05 % of reading + 6.6 mV	
1 kHz to 10 kHz	33 V to 329.999 V	0.08 % of reading + 15 μ V	
10 kHz to 20 kHz	33 V to 329.999 V	0.09 % of reading + 33 μ V	
Equipment to Output AC Voltage At the listed frequencies ^{FO}			
45 Hz to 1 kHz	330 V to 1 020 V	0.05 % of reading + 80 mV	
1 kHz to 5 kHz	330 V to 1 020 V	0.2 % of reading + 100 μ V	Hipot 3565D Internal Procedure: STC-PC-ELE-004
5 kHz to 10 kHz	330 V to 1 020 V	0.2 % of reading + 500 μ V	
Equipment to Output AC Voltage At the listed frequencies ^{FO}			HiPot HD125 Internal Procedure: STC-PC-ELE-004
60 Hz	Up to 5 kV	2 % of reading	
60 Hz	2 kV to 20 kV	2.5 % of reading	HP 3458A Internal Procedure: STC-PC-ELE-006
Equipment to Measure AC Voltage At the listed frequencies ACBAND \leq 2MHz ^{FO}			HP 3458A Internal Procedure: STC-PC-ELE-006
1 Hz to 40 Hz	Up to 10 mV	0.03 % of reading + 0.03 mV	
40 Hz to 1 kHz	Up to 10 mV	0.02 % of reading + 0.011 mV	
45 Hz to 100 kHz	Up to 10 mV	0.09 % of reading + 0.06 mV	
100 kHz to 1 MHz	Up to 10 mV	1.2 % of reading + 0.05 mV	
1 MHz to 4 MHz	Up to 10 mV	7 % of reading + 0.07 mV	
4 MHz to 8 MHz	Up to 10 mV	20 % of reading + 0.08 mV	
Equipment to Measure AC Voltage At the listed frequencies ACBAND \leq 2MHz ^{FO}			
45 Hz to 100 kHz	100 mV to 10 V	0.09 % of reading + 0.06 V	
100 kHz to 1 MHz	100 mV to 10 V	2 % of reading + 0.05 V	
1 MHz to 4 MHz	100 mV to 10 V	4 % of reading + 0.07 V	
4 MHz to 8 MHz	100 mV to 10 V	4 % of reading + 0.08 V	



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Voltage At the listed frequencies ACBAND $\leq 2\text{MHz}^{\text{FO}}$			HP 3458A Internal Procedure: STC-PC-ELE-006
8 MHz to 10 MHz	100 mV to 10 V	15 % of reading + 0.1 V	
45 Hz to 100 kHz	10 V to 100 V	0.12 % of reading + 0.02 V	
45 Hz to 100 kHz	100 V to 1 000 V	0.3 % of reading + 1 V	
Equipment to Measure AC Voltage At the listed frequencies			Vitretek 4620B Internal Procedure STC-PC-ELE-004
20 Hz to 100 Hz	Up to 2 kV	0.07 % of reading + 2 V	
100 Hz to 400 Hz	Up to 2 kV	0.4 % of reading + 4 V	
20 Hz to 100 Hz	2 kV to 20 kV	0.2 % of reading + 20 V	
Equipment to Output AC Current At the listed frequencies $^{\text{FO}}$			Fluke 5500A Intercal Procedure: STC-PC-ELE-006
10 Hz to 20 Hz	0.029 mA to 0.32 999 mA	0.25 % of reading + 0.15 μA	
20 Hz to 45 Hz	0.029 mA to 0.32 999 mA	0.13 % of reading + 0.15 μA	
45 Hz to 1 kHz	0.029 mA to 0.32 999 mA	0.13 % of reading + 0.25 μA	
1 kHz to 5 kHz	0.029 mA to 0.32 999 mA	0.4 % of reading + 0.15 μA	
5 kHz to 10 kHz	0.029 mA to 0.32 999 mA	1.3 % of reading + 0.15 μA	
Equipment to Output AC Current At the listed frequencies			
10 Hz to 20 Hz	0.33 mA to 3.2 999 mA	0.2 % of reading + 0.3 μA	
20 Hz to 45 Hz	0.33 mA to 3.2 999 mA	0.1 % of reading + 0.3 μA	
Equipment to Measure AC Voltage At the listed frequencies ACBAND $\leq 2\text{MHz}^{\text{FO}}$			HP 3458A Intercal Procedure: STC-PC-ELE-006
1 kHz to 20 kHz	Up to 10 mV	0.03 % of reading + 0.011 mV	
20 kHz to 50 kHz	Up to 10 mV	0.1 % of reading + 0.011 mV	
50 kHz to 100 kHz	Up to 10 mV	0.5 % of reading + 0.011 mV	
100 kHz to 300 kHz	Up to 10 mV	4 % of reading + 0.02 mV	



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Voltage At the listed frequencies ACBAND $\leq 2\text{MHz}^{\text{FO}}$			HP 3458A Internal Procedure STC-PC-ELE-006
40 Hz to 1 kHz	100 mV to 10 V	0.007 % of reading + 0.002 V	
1 kHz to 20 kHz	100 mV to 10 V	0.014 % of reading + 0.002 V	
20 kHz to 50 kHz	100 mV to 10 V	0.03 % of reading + 0.002 V	
50 kHz to 100 kHz	100 mV to 10 V	0.08 % of reading + 0.002 V	
100 kHz to 300 kHz	100 mV to 10 V	0.3 % of reading + 0.01 V	
300 kHz to 1 MHz	100 mV to 10 V	1 % of reading + 0.01 V	
1 MHz to 2 MHz	100 mV to 10 V	1.5 % of reading + 0.01 V	
1 Hz to 40 Hz	100 mV to 10 V	0.007 % of reading + 0.004 V	
Equipment to Measure AC Voltage At the listed frequencies ACBAND $\leq 2\text{MHz}^{\text{FO}}$			
1 Hz to 40 Hz	10 V to 100 V	0.02 % of reading + 0.04 V	
40 Hz to 1 kHz	10 V to 100 V	0.02 % of reading + 0.02 V	
1 kHz to 20 kHz	10 V to 100 V	0.02 % of reading + 0.02 V	
20 kHz to 50 kHz	10 V to 100 V	0.035 % of reading + 0.02 V	
50 kHz to 100 kHz	10 V to 100 V	0.12 % of reading + 0.02 V	
100 kHz to 300 kHz	10 V to 100 V	0.4 % of reading + 0.1 V	
300 kHz to 1 MHz	10 V to 100 V	1.5 % of reading + 0.1 V	
Equipment to Measure AC Voltage At the listed frequencies ACBAND $\leq 2\text{MHz}^{\text{FO}}$			
1 Hz to 40 Hz	100 V to 1 000 V	0.04 % of reading + 0.4 V	
40 Hz to 1 kHz	100 V to 1 000 V	0.04 % of reading + 0.2 V	
1 kHz to 20 kHz	100 V to 1 000 V	0.06 % of reading + 0.2 V	
20 kHz to 50 kHz	100 V to 1 000 V	0.12 % of reading + 0.2 V	
50 kHz to 100 kHz	100 V to 1 000 V	0.3 % of reading + 0.2 V	
Equipment to Output AC Current At the listed frequencies $^{\text{FO}}$			Fluke 5500A Internal Procedure STC-PC-ELE-006
45 Hz to 1 kHz	0.33 mA to 3.2 999 mA	0.1 % of reading + 0.3 μA	



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output AC Current At the listed frequencies ^{FO}			Fluke 5500A Internal Procedure STC-PC-ELE-006
1 kHz to 5 kHz	0.33 mA to 3.2 999 mA	0.2 % of reading + 0.3 μ A	
5 kHz to 10 kHz	0.33 mA to 3.2 999 mA	0.6 % of reading + 0.3 μ A	
Equipment to Output AC Current At the listed frequencies ^{FO}			
10 Hz to 20 Hz	3.3 mA to 32. 999 mA	0.2 % of reading + 3 μ A	
20 Hz to 45 Hz	3.3 mA to 32. 999 mA	0.1 % of reading + 3 μ A	
45 Hz to 1 kHz	3.3 mA to 32. 999 mA	0.09 % of reading + 3 μ A	
1 kHz to 5 kHz	3.3 mA to 32. 999 mA	0.2 % of reading + 3 μ A	
5 kHz to 10 kHz	3.3 mA to 32. 999 mA	0.6 % of reading + 3 μ A	
Equipment to Output AC Current At the listed frequencies ^{FO}			
10 Hz to 20 Hz	33 mA to 329. 99 mA	0.2 % of reading + 30 μ A	
20 Hz to 45 Hz	33 mA to 329. 99 mA	0.1 % of reading + 30 μ A	
45 Hz to 1 kHz	33 mA to 329. 99 mA	0.09 % of reading + 30 μ A	
1 kHz to 5 kHz	33 mA to 329. 99 mA	0.2 % of reading + 30 μ A	
5 kHz to 10 kHz	33 mA to 329. 99 mA	0.6 % of reading + 30 μ A	
Equipment to Output AC Current At the listed frequencies ^{FO}			
10 Hz to 45 Hz	0.33 A to 2.19 999 A	0.2 % of reading + 300 μ A	
45 Hz to 1 kHz	0.33 A to 2.19 999 A	0.1 % of reading + 300 μ A	
1 kHz to 5 kHz	0.33 A to 2.19 999 A	0.75 % of reading + 300 μ A	
Equipment to Output AC Current At the listed frequencies ^{FO}			
65 Hz to 500 Hz	2.2 A to 11 A	0.1 % of reading + 2 000 μ A	
500 Hz to 1 kHz	2.2 A to 11 A	0.33 % of reading + 2 000 μ A	
Equipment to Output AC Current At the listed frequencies ^{FO}			Fluke 5500A/ 50 Turn Coil Internal Procedure: STC-PC-ELE-009
65 Hz to 440 Hz	11 A to 100 A	0.5 % of Output + 0.5 A	
45 Hz to 65 Hz	100 A to 500 A	0.5 % of Output + 0.5 A	



Certificate of Accreditation: Supplement
South Texas Calibration Labs /
Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
 Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
 Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Current At the listed frequencies ^{FO}			HP 3458A Internal Procedure STC-PC-ELE-006
10 Hz to 20 Hz	Up to 100 μ A	0.4 % of reading + 0.3 μ A	
20 Hz to 45 Hz	Up to 100 μ A	0.15 % of reading + 0.3 μ A	
45 Hz to 100 Hz	Up to 100 μ A	0.06 % of reading + 0.3 μ A	
100 Hz to 5 kHz	Up to 100 μ A	0.06 % of reading + 0.3 μ A	
Equipment to Measure AC Current At the listed frequencies ^{FO}			
10 Hz to 20 Hz	1 mA to 100 mA	0.4 % of reading + 0.2 mA	
20 Hz to 45 Hz	1 mA to 100 mA	0.15 % of reading + 0.2 mA	
45 Hz to 100 Hz	1 mA to 100 mA	0.06 % of reading + 0.2 mA	
100 Hz to 5 kHz	1 mA to 100 mA	0.03 % of reading + 0.2 mA	
5 kHz to 20 kHz	1 mA to 100 mA	0.06 % of reading + 0.2 mA	
Equipment to Measure AC Current At the listed frequencies ^{FO}			
20 kHz to 50 kHz	1 mA to 100 mA	0.4 % of reading + 0.4 mA	
50 kHz to 100 kHz	1 mA to 100 mA	0.55 % of reading + 1.5 mA	
Equipment to Measure AC Current At the listed frequencies ^{FO}			
10 Hz to 20 Hz	100 mA to 1 A	0.4 % of reading + 0.2 mA	
20 Hz to 45 Hz	100 mA to 1 A	0.16 % of reading + 0.2 mA	
45 Hz to 100 Hz	100 mA to 1 A	0.08 % of reading + 0.2 mA	
100 Hz to 5 kHz	100 mA to 1 A	0.1 % of reading + 0.2 mA	
5 kHz to 20 kHz	100 mA to 1 A	0.3 % of reading + 0.2 mA	
20 kHz to 50 kHz	100 mA to 1 A	1 % of reading + 0.4 mA	
Equipment to Output Resistance ^{FO}			
	33 k Ω to 110 k Ω	0.011 % of reading + 6 Ω	
	110 k Ω to 330 k Ω	0.012 % of reading + 6 Ω	
	11 Ω to 33 Ω	0.012 % of reading + 15 m Ω	
	33 Ω to 110 Ω	0.009 % of reading + 15 m Ω	
	110 Ω to 330 Ω	0.009 % of reading + 15 m Ω	
	330 Ω to 1.1 k Ω	0.009 % of reading + 60 m Ω	
	1.1 k Ω to 3.3 k Ω	0.009 % of reading + 60 m Ω	



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output Resistance ^{FO}	3.3 k Ω to 11 k Ω	0.009 % of reading + 0.6 Ω	HP 3458A Internal Procedure STC-PC-ELE-006
Equipment to Output Resistance ^{FO}	11 k Ω to 33 k Ω	0.009 % of reading + 0.6 Ω	Fluke 5500A Internal Procedure: STC-PC-ELE-006
	1.1 M Ω to 3.3 M Ω	0.015 % of reading + 55 Ω	
	3.3 M Ω to 11 M Ω	0.05 % of reading + 550 Ω	
	11 M Ω to 33 M Ω	0.1 % of reading + 550 Ω	
	33 M Ω to 110 M Ω	0.5 % of reading + 5.5 k Ω	
	110 M Ω to 330 M Ω	0.5 % of reading + 16.5 k Ω	
	330 k Ω to 1.1 M Ω	0.015 % of reading + 55 Ω	
Equipment to Measure Resistance ^{FO}	Up to 10 Ω	0.001 5 % of reading + 50 $\mu\Omega$	HP 3458A Internal Procedure STC-PC-ELE-006
	Up to 100 Ω	0.001 2 % of reading + 500 $\mu\Omega$	
	Up to 1 k Ω	0.001 % of reading + 500 $\mu\Omega$	
	Up to 10 k Ω	0.001 % of reading + 5 m Ω	
	Up to 100 k Ω	0.001 % of reading + 50 m Ω	
	Up to 1 M Ω	0.001 5 % of reading + 2 Ω	
	Up to 10 M Ω	0.005 % of reading + 100 Ω	
	Up to 100 M Ω	0.05 % of reading + 1 k Ω	
	Up to 1 G Ω	0.5 % of reading + 10 k Ω	
Equipment to Output Capacitance At the listed frequencies ^{FO}			Fluke 5500A Internal Procedure: STC-PC-ELE-006
50 Hz to 1 000 Hz	0.33 η F to 10.999 η F	0.5 % of reading + 0.01 η F	
50 Hz to 1 000 Hz	11 η F to 109 η F	0.25 % of reading + 0.1 η F	
50 Hz to 1 000 Hz	110 η F to 329.99 η F	0.25 % of reading + 0.3 η F	
50 Hz to 1 000 Hz	0.33 μ F to 1.099 9 μ F	0.25 % of reading + 1 η F	
50 Hz to 1 000 Hz	1.1 μ F to 3.299 9 μ F	0.35 % of reading + 3 η F	
50 Hz to 400 Hz	3.3 μ F to 10.999 μ F	0.35 % of reading + 10 η F	
50 Hz to 400 Hz	11 μ F to 32.999 μ F	0.4 % of reading + 30 η F	
50 Hz to 200 Hz	33 μ F to 109.99 μ F	0.5 % of reading + 10 η F	
50 Hz to 100 Hz	11 μ F to 329.99 μ F	0.7 % of reading + 300 η F	
50 Hz to 100 Hz	0.33 mF to 1.1 mF	1 % of reading + 300 η F	



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Output Capacitance At the listed frequencies ^{FO}			Capacitance Decade CS-301 Internal Procedure STC-PC-ELE-003
120 Hz	0.2 η F to 200 μ F	1 % of reading + 3 pF	
1 kHz	2 η F to 2 000 μ F	1 % of reading + 3 pF	
Equipment to Measure Capacitance At the listed frequencies ^{FO}			GenRad 1658 Internal Procedure STC-PC-ELE-003
1 kHz	0.2 η F to 200 μ F	0.1 % of reading	
120 Hz	2 η F to 2 000 μ F	0.1 % of reading	
Equipment to Output Inductance At the listed frequencies ^{FO}			Inductance Decade LS-400 Internal Procedure STC-PC-ELE-003
300 Hz to 200 kHz	0.1 H to 0.999 H	2 % of reading + 0.5 μ H	
Equipment to Measure Inductance At the listed frequencies ^{FO}			GenRad 1658 Internal Procedure: STC-PC-ELE-003
1 kHz	0.2 mH to 200 H	0.1 % of reading	
120 kHz	2 mH to 2 000 H	0.1 % of reading	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type B ^{FO}	600 °C to 800 °C	0.44 °C	Fluke 5500A Electrical Simulation of Thermocouple Output Internal Procedure: STC-PC-TER-002
	800 °C to 1 000 °C	0.34 °C	
	1 000 °C to 1 500 °C	0.3 °C	
	1 500 °C to 1 820 °C	0.33 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type C ^{FO}	0 °C to 150 °C	0.3 °C	
	150 °C to 650 °C	0.26 °C	
	650 °C to 1 000 °C	0.31 °C	
	1 000 °C to 1 800 °C	0.5 °C	
	1 800 °C to 2 316 °C	0.84 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type E ^{FO}	-250 °C to -100 °C	0.5 °C	
	-25 °C to 350 °C	0.14 °C	
	350 °C to 650 °C	0.16 °C	
	650 °C to 1 000 °C	0.21 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type J	-100 °C to -25 °C	0.16 °C	
	-210 °C to -100 °C	0.27 °C	



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type J ^{FO}	-100 °C to -30 °C	0.16 °C	Fluke 5500A Electrical Simulation of Thermocouple Output Internal Procedure STC-PC-TER-002
	-30 °C to 150 °C	0.14 °C	
	150 °C to 760 °C	0.17 °C	
	760 °C to 1 200 °C	0.23 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type K ^{FO}	-200 °C to -100 °C	0.33 °C	
	-100 °C to -25 °C	0.18 °C	
	-25 °C to 120 °C	0.16 °C	
	120 °C to 1 000 °C	0.26 °C	
	1 000 °C to 1 372 °C	0.4 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type L ^{FO}	-200 °C to -100 °C	0.37 °C	
	-100 °C to 800 °C	0.26 °C	
	800 °C to 900 °C	0.17 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type N ^{FO}	-200 °C to -100 °C	0.4 °C	
	-100 °C to -25 °C	0.22 °C	
	-25 °C to 120 °C	0.19 °C	
	120 °C to 410 °C	0.18 °C	
	410 °C to 1 300 °C	0.27 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type R ^{FO}	0 °C to 250 °C	0.57 °C	Fluke 5500A Electrical Simulation of Thermocouple Output Internal Procedure STC-PC-TER-002
	250 °C to 400 °C	0.35 °C	
	400 °C to 1 000 °C	0.33 °C	
	1 000 °C to 1 767 °C	0.4 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type S ^{FO}	0 °C to 250 °C	0.47 °C	
	250 °C to 1 000 °C	0.36 °C	
	1 000 °C to 1 400 °C	0.37 °C	
	1 400 °C to 1 767 °C	0.46 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type T ^{FO}	-250 °C to -150 °C	0.63 °C	
	-150 °C to 0 °C	0.24 °C	
	0 °C to 120 °C	0.16 °C	
	120 °C to 400 °C	0.14 °C	
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type U ^{FO}	-200 °C to 0 °C	0.56 °C	
	0 °C to 600 °C	0.27 °C	



Certificate of Accreditation: Supplement
ISO/IEC 17025:2005

South Texas Calibration Labs
Morelos # 130, Col. Centro
Cd. Gustavo Díaz Ordaz, Tamaulipas 88400
Contact: Fernando Casiano. Phone: (891) 938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Calibration, Indication and Control Equipment used with RTD Pt 385, 100 Ω^{FO}	200 °C to -80 °C	0.05 °C	Fluke 5500A Electrical Simulation of RTD Output Internal Procedure STC-PC-TER-002
	-80 °C to 0 °C	0.05 °C	
	0 °C to 100 °C	0.07 °C	
	100 °C to 300 °C	0.09 °C	
	300 °C to 400 °C	0.1 °C	
	400 °C to 630 °C	0.12 °C	
	630 °C to 800 °C	0.23 °C	
Temperature Calibration, Indication and Control Equipment used with RTD Pt 385, 200 Ω^{FO}	-200 °C to -80 °C	0.04 °C	
	-80 °C to 0 °C	0.04 °C	
	0 °C to 100 °C	0.04 °C	
	100 °C to 260 °C	0.05 °C	
	260 °C to 300 °C	0.12 °C	
	300 °C to 400 °C	0.13 °C	
	400 °C to 600 °C	0.14 °C	
Temperature Calibration, Indication and Control Equipment used with RTD Pt 385, 500 Ω^{FO}	-200 °C to -80 °C	0.04 °C	
	-80 °C to 0 °C	0.05 °C	
	0 °C to 100 °C	0.05 °C	
	100 °C to 260 °C	0.06 °C	
	260 °C to 300 °C	0.08 °C	
	300 °C to 400 °C	0.08 °C	
	400 °C to 600 °C	0.09 °C	
Temperature Calibration, Indication and Control Equipment used with RTD Pt 385, 1 000 Ω^{FO}	-200 °C to -80 °C	0.03 °C	
	-80 °C to 0 °C	0.03 °C	
	0 °C to 100 °C	0.04 °C	
	100 °C to 260 °C	0.05 °C	
	260 °C to 300 °C	0.06 °C	
	300 °C to 400 °C	0.07 °C	
	400 °C to 600 °C	0.07 °C	
	600 °C to 630 °C	0.23 °C	



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Controllers ^{FO}	-20 °C to 150 °C	0.8 °C	Fluke 5500A Internal Procedure STC-PC-TER-002

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Chambers Recorders Thermometers ^F	-10 °C to 65 °C	0.19 °C	Omega OM-DVTH Comparison Internal Procedure: STC-PC-TER-003
Soldering Station ^O	0 °C to 700 °C	0.45 °C	Hakko FG-101 Comparison Internal Procedure: STC-PC-TER-003
Hygrometers, Recorders and Chambers ^F	10 % RH to 95 % RH	1.2 % of reading	Omega OM-DVTH Comparison Internal Procedure: STC-PC-TER-003
Laser Thermometers (IR) ^{FO} Fixed Point	50 °C 100 °C 150 °C	0.18 °C 0.18 °C 0.18 °C	Fluke 9135 Internal Procedure: STC-PC-TER-003

Chemical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
pH Meter ^{FO}	4.01 pH	0.027 pH	Omega PHH-60BMS Comparison Internal Procedure: STC-PC-QUI-001 STC-PC-QUI-002
	7.01 pH	0.027 pH	
	10.01 pH	0.027 pH	
Conductivity Meter ^{FO}	45 μ S/cm	0.61 μ S/cm	
	450 μ S/cm	5.7 μ S/cm	
	1 500 μ S/cm	8.7 μ S/cm	
	4 500 μ S/cm	26 μ S/cm	



Certificate of Accreditation: Supplement

South Texas Calibration Labs / Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pressure Gage ^{FO}	3 psi to 30 psi	0.1 % of reading	Ametek CPC030CINDG ASME B40.100
	0.01 psi to 100 psi	0.2 % of reading	Ashcroft 25D1005PS02L100 ASME B40.100
	0.1 psi to 1 000 psi	0.25 % of reading	
	1 psi to 10 000 psi	0.5 % of reading	Druck DPI104-2- 10000PSI-SG ASME B40.100
Vacuum Gage ^{FO}	-10.7 psi to 0 psi (-101.35 kPa to 0 kPa)	0.05 % of reading	Ametek CPC030CINDG ASME B40 BS EN 837-1
Torque (Analyzer, Wrench, Transducer, Screwdriver) ^F	5 lbf·in to 100 lbf·in (0.011 N·m to 11.29 N·m)	0.63 lbf·in (2.802 38 N·m)	Torque Analyzer ISO 6789
	1 lbf·in to 7 200 lbf·in (0.11 N·m to 813.4 N·m)	0.5 % of reading	Weight Set Class F NIST (105-1)
	1 lbf·ft to 600 lbf·ft (1.35 N·m to 813.4 N·m)	0.55 % of reading	
Direct Verification of Durometer Hardness Tester Types: A, B, C, D, E, DO, O, M Extension at Zero reading	2.46 mm to 2.54 mm	4.5 µm	ASTMD-2240 Video Measuring Machine
Indenter Shape (Not all parameters apply to all Durometer Types)			
Indenter Diameter		4.5 µm	
Indenter Tip Diameter		4.5 µm	
Indenter Tip Radius		4.5 µm	
Indenter Tip Angle ^F		0.1°	
Verification of Durometer Spring Type A, B, E & O ^F	0.55 N to 8.05 N	1.2 N	ASTMD-2240 Analytical Balance
Verification of Durometer Spring Type C, D & DO ^F	4.445 N to 44.45 N	0.8 N	



Certificate of Accreditation: Supplement
South Texas Calibration Labs /
Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
 Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
 Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Indirect Verification of Rockwell Hardness Testers HRB ^O	40 HRB to 59 HRB	0.36 HRB	Test Block ASTM E18-08b
	60 HRB to 89 HRB	0.34 HRB	
	90 HRB to 100 HRB	0.26 HRB	
Indirect Verification of Rockwell Hardness Testers HRC ^O	20 HRC to 39 HRC	0.21 HRC	
	40 HRC to 59 HRC	0.21 HRC	
	60 HRC to 70 HRC	0.21 HRC	
Leak Tester ^{FO}	Up to 20 sccm	0.12 sccm	Intertech CM-15 Internal Procedure: STC-PC-MEC-009
Relief Valves ^{FO}	Up to 10 000 psi	3.6 psi	Dead Weight Tester NOM-093-SCFI

Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USE
Balance and Scale ^O	1 g to 500 g (Res.= 0.01 mg)	$(1 \times 10^{-4} + 2.7 \times 10^{-6}Wt) \text{ g}$	Weights Class 1 SIM MWG7/cg-01 Internal Procedure: STC-PC-MAS-002
	500 g to 5 kg (Res.= 0.1 g)	$(1.2 \times 10^{-1} + 7.8 \times 10^{-7}Wt) \text{ g}$	
	5 kg to 100 kg (Res.= 1 g)	$(1.1 + 1.5 \times 10^{-5}Wt) \text{ g}$	Weights Class 4 SIM MWG7/cg-01 Internal Procedure: STC-PC-MAS-002
	1 lb to 60 lb (Res.= 0.000 1 lb)	$(2 \times 10^{-4} + 1.5 \times 10^{-5}Wt) \text{ lb}$	
	60 lb to 5 000 lb (Res.= 1 lb)	$(1.2 + 1 \times 10^{-7}Wt) \text{ lb}$	
Weights Class 3 ^F	1 g	0.03 mg	Master Weights ABBA Method Class 2
	2 g	0.04 mg	
	5 g	0.046 mg	
Weights Class 3 ^F	10 g	0.048 mg	
	20 g	0.08 mg	



Certificate of Accreditation: Supplement
South Texas Calibration Labs /
Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
 Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
 Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	BEST MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USE
Weights Class 3 ^F	50 g	0.14 mg	Master Weights ABBA Method Class 2
	100 g	0.29 mg	
	200 g	0.57 mg	
	500 g	1.4 mg	
Weights Class 4 ^F	1 kg	5.8 mg	
	2 kg	12 mg	
Weights Class 6 ^F	5 kg	140 mg	Master Weights ABBA Method Class 4
	10 kg	280 mg	
	20 kg	590 mg	
Weights Class F ^F	1 lb	42 μ lb	
	2 lb	59 μ lb	
	5 lb	150 μ lb	
	10 lb	320 μ lb	
	20 lb	610 μ lb	
	50 lb	1 500 μ lb	
Force-Compression and Tension ^{FO}	Up to 10 000 lbf	0.6 % of reading	Load Cell and Weights Comparison LC101-10K ASTM 4-16

Acoustic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Sound Level Meter ^{FO}	94 dB to 114 dB	0.28 dB	Sound Level Calibrator OIML R102

Time and Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Frequency – Source ^{FO}	0.01 Hz to 11.999 kHz	0.002 5 % of reading + 1 MHz	Fluke 5500A Internal Procedure STC-PC-TYF-002
	2 kHz to 2 MHz	0.002 5 % of reading + 15 MHz	
Frequency – Measure ^{FO}	1 Hz to 40 Hz	0.05 % of reading	Agilent 53132A Internal Procedure STC-PC-TYF-005
	40 Hz to 10 MHz	0.01 % of reading	



Certificate of Accreditation: Supplement
South Texas Calibration Labs /
Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
 Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
 Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

Time and Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
StopWatch / Timer ^{FO}	2 s to 86 400 s	35 ms	960-12 Totalize Method Internal Procedure: STC-PC-TYF-001
Tachometer ^{FO}	1 Rad/s to 3 141 Rad/s	0.2 Rad/s	Monarch Nova-Strobe Internal Procedure STC-PC-ELE-008
Radio Frequency			Agilent E4433B
Equipment to Output Frequency ^{FO}	250 kHz to 4 GHz	7.6×10^{-2} kHz	Internal Procedure STC-PC-TYF-002
Equipment to Measure Frequency ^{FO}	10 Hz to 5 GHz	0.008 2 Hz	Agilent 53132A Internal Procedure STC-PC-TYF-005
Equipment to Output Power ^{FO}	3 μ W to 100 mW	0.007 6 μ W	HP 11683A Internal Procedure STC-PC-TYF-004
	20 dB to -136 dB	0.062 dB	Agilent E4433B Internal Procedure STC-PC-TYF-004
Equipment to Measure Power ^{FO}	30 dB to -120 dB	0.019 dB	HP 8902A / 11722A Internal Procedure STC-PC-TYF-005
Audio Frequency ^{FO}	1 Hz to 250 kHz	5.8×10^{-2} Hz	
Audio Distortion Measure ^{FO}	Up to 99 dB	0.014 % of reading	HP 8903A Internal Procedure STC-PC-TYF-005
Audio RMS Measure ^{FO}	100 mVrms to 7 Vrms	5.8×10^{-5} Vrms	Agilent 34401A Internal Procedure STC-PC-TYF-005
VPP ^{FO}	0.2 mVpp to 100 Vpp	1.4 mVpp	PG506 Internal Procedure STC-PC-TYF-003

- The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.



Certificate of Accreditation: Supplement
South Texas Calibration Labs /
Fernando Damian Casiano Flores

Morelos #130, Colonia Centro
Ciudad Gustavo Díaz Ordaz, Tamaulipas, México. C.P. 88400
Contact: Fernando Casiano Flores Phone: 891-938-3738

Accreditation is granted to the facility to perform the following calibrations:

2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the Procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the
4. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
5. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer^O would mean that the laboratory performs this calibration onsite at the customer's location.
6. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
7. The term L represents length in inches or millimeters as appropriate to the uncertainty statement.
8. The term Wt. represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.
9. This is the primary site for all quality management system activities.